

Instrument name: DB1			11/02/2015 07:41:16		Page 1/1
Start time: 10/05/2015 11:08:00		Minimum	Maximum	Mean value	Limit values
End time: 11/02/2015 07:08:00	Temperature [°F]	68.00	74.30	71.20	50.0/100.0
Measurement channels: 2	Humidity [%rH]	-1.10	42.10	-0.45	0.0/100.0
Measured values: 670					
36685025					

