

|                                 |                  |         |                     |            |              |      |     |
|---------------------------------|------------------|---------|---------------------|------------|--------------|------|-----|
| Instrument name: DB1            |                  |         | 03/01/2016 07:43:08 |            |              | Page | 1/1 |
| Start time: 02/02/2016 08:20:00 |                  | Minimum | Maximum             | Mean value | Limit values |      |     |
| End time: 03/01/2016 07:20:00   | Temperature [°F] | 66.92   | 80.60               | 71.46      | 50.0/100.0   |      |     |
| Measurement channels: 2         | Humidity [%rH]   | -1.20   | 27.70               | -0.78      | 0.0/100.0    |      |     |
| Measured values: 672            |                  |         |                     |            |              |      |     |
| 36685025                        |                  |         |                     |            |              |      |     |
|                                 |                  |         |                     |            |              |      |     |
|                                 |                  |         |                     |            |              |      |     |
|                                 |                  |         |                     |            |              |      |     |

