


09-06-2016 10:38:07



LHO Vertex RGA while isolated from the Corner Station, pumped by the locally mounted turbo (with an air leak confirmed at the vent valve) and with the locally mounted 3.3×10^{-8} torr*L/sec Krypton leak open. SEM = 1500 volts.