
LHOX-End RxPD and TxPD Calibration Trends

Table of Contents

About	1
WS/TX Ratio when WS is at TX (Inner Beam)	2
WS/TX Ratio when WS is at TX (Outer Beam)	3
WS/TX Ratio when WS is at RX (Inner Beam)	4
WS/TX Ratio when WS is at RX (Outer Beam)	5
RX/TX Ratio (Inner Beam)	6
RX/TX Ratio (Outer Beam)	7
Optical Efficiency of Inner Beam	8
Optical Efficiency of Outer Beam	9
Total Optical Efficiency	11
TX/WS Ratio	12
RX/WS Ratio	13
TX/WS Ratio (Corrected for OE)	14
RX/WS Ratio (Corrected for OE)	15
Power Imbalance	16
Summary	17

About

This document contains the Pcal PD (TxPD and RxPD) Calibration trends. The first six sections contain the six ratios measured at the end-station labeled as m1, m2m6. The section that follows contains the relevant information calculated from these measurements which include Optical Efficiency, Power Imbalance, TX/WS and RX/WS ratio.

Understanding Each Section

Each section contains a list of measurements with Magnitude, Standard Error (Std Err) and Relative Error (Rel Err) for each measurement. The list is followed by two plot figures with Magnitude on the first plot and the the Normalized Magnitude on the second. Each section ends with a summary that contains the weighted mean of all the measurement along with their Standard Deviation (Std Dev), Std Err and Rel Err where each of


$$\begin{aligned} \text{Mean} &= \text{sum}(x(i)*w(i))/\text{sum}(w(i)) \\ \text{Std Dev} &= \text{sqrt}(\text{sum}(w(i)*(x(i)-x_mean)^2)/((n-1)/n*\text{sum}(w(i)))) \\ \text{Std Err} &= \text{Std Dev}/\text{sqrt}(n) \\ \text{Rel Err} &= \text{Std Err}/\text{Mean} \end{aligned}$$

Report created on 08-Nov-2016

WS/TX Ratio when WS is at TX (Inner Beam)

List of Measurements

<i>Date</i>	<i>$m1 \pm SE_{\{m1\}}$</i>	<i>Normalized</i>
<i>D20150804</i>	<i>-0.158554 ± 0.000007</i>	<i>(1 ± 0.000042)</i>
<i>D20150827</i>	<i>-0.158839 ± 0.000004</i>	<i>(1 ± 0.000028)</i>
<i>D20151013</i>	<i>-0.158837 ± 0.000005</i>	<i>(1 ± 0.000032)</i>
<i>D20160108</i>	<i>-0.157751 ± 0.000005</i>	<i>(1 ± 0.000035)</i>
<i>D20160505</i>	<i>-0.158272 ± 0.000005</i>	<i>(1 ± 0.000034)</i>
<i>D20160628</i>	<i>-0.158053 ± 0.000005</i>	<i>(1 ± 0.000033)</i>
<i>D20160922</i>	<i>-0.157555 ± 0.000009</i>	<i>(1 ± 0.000060)</i>
<i>D20161108</i>	<i>-0.141888 ± 0.000005</i>	<i>(1 ± 0.000035)</i>




Summary of WS/TX when WS is at TX (Inner beam) [m1]:
Mean value: -0.155860
Standard deviation: 0.006355

Standard Error: 0.002396
Relative Standard Error: 0.015374

WS/TX Ratio when WS is at TX (Outer Beam)

List of Measurements

Date	$m2 \pm SE_{\{m2\}}$	Normalized
D20150804	-0.158016 ± 0.000004	(1 ± 0.000027)
D20150827	-0.158288 ± 0.000004	(1 ± 0.000028)
D20151013	-0.158597 ± 0.000005	(1 ± 0.000030)
D20160108	-0.159262 ± 0.000005	(1 ± 0.000029)
D20160505	-0.158328 ± 0.000005	(1 ± 0.000033)
D20160628	-0.157681 ± 0.000005	(1 ± 0.000032)
D20160922	-0.158094 ± 0.000010	(1 ± 0.000062)
D20161108	-0.142053 ± 0.000005	(1 ± 0.000034)




Summary of WS/TX when WS is at TX (Outer Beam) [m2]:

Mean value: -0.156248
Standard deviation: 0.005864
Standard Error: 0.002211
Relative Standard Error: 0.014151

WS/TX Ratio when WS is at RX (Inner Beam)

List of Measurements

<i>Date</i>	<i>m3 ± SE_{m3}</i>	<i>Normalized</i>
<i>D20150804</i>	<i>-0.156120 ± 0.000006</i>	<i>(1 ± 0.000039)</i>
<i>D20150827</i>	<i>-0.156101 ± 0.000005</i>	<i>(1 ± 0.000032)</i>
<i>D20151013</i>	<i>-0.154453 ± 0.000005</i>	<i>(1 ± 0.000034)</i>
<i>D20160108</i>	<i>-0.141148 ± 0.000008</i>	<i>(1 ± 0.000057)</i>
<i>D20160505</i>	<i>-0.132887 ± 0.000011</i>	<i>(1 ± 0.000081)</i>
<i>D20160628</i>	<i>-0.123153 ± 0.000009</i>	<i>(1 ± 0.000073)</i>
<i>D20160922</i>	<i>-0.155734 ± 0.000009</i>	<i>(1 ± 0.000060)</i>
<i>D20161108</i>	<i>-0.140186 ± 0.000004</i>	<i>(1 ± 0.000032)</i>



Summary of WS/TX when WS is at RX (Inner Beam) [m3]:
Mean value: -0.147657
Standard deviation: 0.010655
Standard Error: 0.004018
Relative Standard Error: 0.027211

WS/TX Ratio when WS is at RX (Outer Beam)


List of Measurements

Date	$m4 \pm SE_{\{m4\}}$	Normalized
D20150804	-0.154938 ± 0.000006	(1 ± 0.000038)
D20150827	-0.153203 ± 0.000005	(1 ± 0.000031)
D20151013	-0.140688 ± 0.000010	(1 ± 0.000068)
D20160108	-0.107301 ± 0.000014	(1 ± 0.000127)
D20160505	-0.096847 ± 0.000018	(1 ± 0.000181)
D20160628	-0.072041 ± 0.000018	(1 ± 0.000253)
D20160922	-0.156012 ± 0.000009	(1 ± 0.000061)

D20161108

-0.139772 ± 0.000005

(1 ± 0.000036)



Summary of WS/TX when WS is at RX (Outer Beam) [m4]:


Mean value: -0.144714
Standard deviation: 0.017560
Standard Error: 0.006621
Relative Standard Error: 0.045755

RX/TX Ratio (Inner Beam)

List of Measurements

Date	$m5 \pm SE_{\{m5\}}$	Normalized
D20150804	0.628735 ± 0.000004	(1 ± 0.000006)
D20150827	0.627025 ± 0.000003	(1 ± 0.000005)
D20151013	0.620693 ± 0.000007	(1 ± 0.000011)
D20160108	0.570708 ± 0.000023	(1 ± 0.000040)
D20160505	0.533749 ± 0.000043	(1 ± 0.000080)

<i>D20160628</i>	<i>0.487934 ± 0.000054</i>	<i>(1 ± 0.000112)</i>
<i>D20160922</i>	<i>0.627214 ± 0.000006</i>	<i>(1 ± 0.000010)</i>
<i>D20161108</i>	<i>0.628219 ± 0.000003</i>	<i>(1 ± 0.000005)</i>



Summary of RX/TX (Inner Beam) [m5]:


<i>Mean value:</i>	<i>0.626691</i>
<i>Standard deviation:</i>	<i>0.008198</i>
<i>Standard Error:</i>	<i>0.003091</i>
<i>Relative Standard Error:</i>	<i>0.004933</i>

RX/TX Ratio (Outer Beam)

List of Measurements

<i>Date</i>	<i>m6 ± SE_{m6}</i>	<i>Normalized</i>
<i>D20150804</i>	<i>0.623505 ± 0.000005</i>	<i>(1 ± 0.000008)</i>
<i>D20150827</i>	<i>0.612750 ± 0.000015</i>	<i>(1 ± 0.000024)</i>
<i>D20151013</i>	<i>0.563327 ± 0.000033</i>	<i>(1 ± 0.000059)</i>

<i>D20160108</i>	<i>0.429049 ± 0.000052</i>	<i>(1 ± 0.000122)</i>
<i>D20160505</i>	<i>0.382774 ± 0.000063</i>	<i>(1 ± 0.000164)</i>
<i>D20160628</i>	<i>0.278553 ± 0.000088</i>	<i>(1 ± 0.000315)</i>
<i>D20160922</i>	<i>0.628100 ± 0.000003</i>	<i>(1 ± 0.000005)</i>
<i>D20161108</i>	<i>0.626777 ± 0.000003</i>	<i>(1 ± 0.000005)</i>



Summary of RX/TX (Outer Beam) [m6]:


<i>Mean value:</i>	<i>0.625788</i>
<i>Standard deviation:</i>	<i>0.014623</i>
<i>Standard Error:</i>	<i>0.005514</i>
<i>Relative Standard Error:</i>	<i>0.008811</i>

Optical Efficiency of Inner Beam

List of Measurements

<i>Date</i>	<i>e_i ± SE_{e_i}</i>	<i>Normalized</i>
<i>D20150804</i>	<i>0.984647 ± 0.000058</i>	<i>(1 ± 0.000058)</i>

D20150827	0.982762 ± 0.000043	(1 ± 0.000043)
D20151013	0.972403 ± 0.000046	(1 ± 0.000048)
D20160108	0.894753 ± 0.000066	(1 ± 0.000074)
D20160505	0.839608 ± 0.000088	(1 ± 0.000105)
D20160628	0.779192 ± 0.000080	(1 ± 0.000103)
D20160922	0.988443 ± 0.000085	(1 ± 0.000086)
D20161108	0.988001 ± 0.000047	(1 ± 0.000048)



Summary of Optical Efficiency (Inner Beam):


Mean value:	0.953230
Standard deviation:	0.064648
Standard Error:	0.024377
Relative Standard Error:	0.025573

Optical Efficiency of Outer Beam

List of Measurements

LHOX-End RxPD and Tx-
PD Calibration Trends

<i>Date</i>	<i>e_o ± SE_{e_o}</i>	<i>Normalized</i>
D20150804	0.980522 ± 0.000046	(1 ± 0.000047)
D20150827	0.967879 ± 0.000042	(1 ± 0.000043)
D20151013	0.887079 ± 0.000074	(1 ± 0.000083)
D20160108	0.673740 ± 0.000130	(1 ± 0.000193)
D20160505	0.611689 ± 0.000184	(1 ± 0.000301)
D20160628	0.456880 ± 0.000255	(1 ± 0.000557)
D20160922	0.986835 ± 0.000087	(1 ± 0.000088)
D20161108	0.983944 ± 0.000050	(1 ± 0.000051)




Summary of Optical Efficiency (Outer Beam):

Mean value: 0.948670
 Standard deviation: 0.091498
 Standard Error: 0.034502
 Relative Standard Error: 0.036368

Total Optical Efficiency

List of Measurements

<i>Date</i>	<i>$e \pm SE_{\{e\}}$</i>	<i>Normalized</i>
<i>D20150804</i>	<i>0.982588 ± 0.000036</i>	<i>(1 ± 0.000037)</i>
<i>D20150827</i>	<i>0.975333 ± 0.000029</i>	<i>(1 ± 0.000030)</i>
<i>D20151013</i>	<i>0.929774 ± 0.000040</i>	<i>(1 ± 0.000043)</i>
<i>D20160108</i>	<i>0.783720 ± 0.000053</i>	<i>(1 ± 0.000067)</i>
<i>D20160505</i>	<i>0.725628 ± 0.000067</i>	<i>(1 ± 0.000093)</i>
<i>D20160628</i>	<i>0.618226 ± 0.000066</i>	<i>(1 ± 0.000106)</i>
<i>D20160922</i>	<i>0.987638 ± 0.000060</i>	<i>(1 ± 0.000061)</i>
<i>D20161108</i>	<i>0.985971 ± 0.000034</i>	<i>(1 ± 0.000034)</i>



Summary of Total Optical Efficiency:


Mean value: 0.927445
Standard deviation: 0.109459

Standard Error: 0.041274
Relative Standard Error: 0.044503

TX/WS Ratio

List of Measurements

Date	$R_{TW} \pm SE_{\{R_{TW}\}}$	Normalized
D20150804	-3.158856 ± 0.000008	(1 ± 0.000003)
D20150827	-3.153314 ± 0.000006	(1 ± 0.000002)
D20151013	-3.150260 ± 0.000007	(1 ± 0.000002)
D20160108	-3.154444 ± 0.000007	(1 ± 0.000002)
D20160505	-3.158561 ± 0.000008	(1 ± 0.000002)
D20160628	-3.167227 ± 0.000007	(1 ± 0.000002)
D20160922	-3.168082 ± 0.000014	(1 ± 0.000004)
D20161108	-3.521857 ± 0.000007	(1 ± 0.000002)




Summary of TX/WS Ratio:

Mean value: -3.209675
Standard deviation: 0.136908
Standard Error: 0.051624
Relative Standard Error: 0.016084

RX/WS Ratio

List of Measurements

<i>Date</i>	<i>R_RW ± SE_{R_RW}</i>	<i>Normalized</i>
<i>D20150804</i>	<i>-4.025735 ± 0.000111</i>	<i>(1 ± 0.000028)</i>
<i>D20150827</i>	<i>-4.008193 ± 0.000102</i>	<i>(1 ± 0.000025)</i>
<i>D20151013</i>	<i>-4.011358 ± 0.000194</i>	<i>(1 ± 0.000048)</i>
<i>D20160108</i>	<i>-4.020935 ± 0.000378</i>	<i>(1 ± 0.000094)</i>
<i>D20160505</i>	<i>-3.984463 ± 0.000534</i>	<i>(1 ± 0.000134)</i>
<i>D20160628</i>	<i>-3.914288 ± 0.000823</i>	<i>(1 ± 0.000210)</i>
<i>D20160922</i>	<i>-4.026718 ± 0.000174</i>	<i>(1 ± 0.000043)</i>
<i>D20161108</i>	<i>-4.482806 ± 0.000110</i>	<i>(1 ± 0.000024)</i>



Summary of RX/WS Ratio:


Mean value: -4.132950
Standard deviation: 0.216386
Standard Error: 0.081594
Relative Standard Error: 0.019742

TX/WS Ratio (Corrected for OE)

List of Measurements

Date	$R_Twc \pm SE_{\{R_Twc\}}$	Normalized
D20150804	-3.186598 ± 0.000059	(1 ± 0.000018)
D20150827	-3.192691 ± 0.000047	(1 ± 0.000015)
D20151013	-3.264901 ± 0.000069	(1 ± 0.000021)
D20160108	-3.536928 ± 0.000108	(1 ± 0.000031)
D20160505	-3.660766 ± 0.000147	(1 ± 0.000040)
D20160628	-3.914444 ± 0.000165	(1 ± 0.000042)
D20160922	-3.187786 ± 0.000097	(1 ± 0.000030)

D20161108 -3.546735 ± 0.000061 (1 ± 0.000017)



Summary of TX/WS Ratio (Corrected for OE):


Mean value: -3.317500
Standard deviation: 0.197946
Standard Error: 0.074641
Relative Standard Error: 0.022499

RX/WS Ratio (Corrected for OE)

List of Measurements

Date	$R_Rwc \pm SE_{\{R_Rwc\}}$	Normalized
D20150804	-3.990379 ± 0.000076	(1 ± 0.000019)
D20150827	-3.958141 ± 0.000077	(1 ± 0.000019)
D20151013	-3.865381 ± 0.000143	(1 ± 0.000037)
D20160108	-3.533387 ± 0.000260	(1 ± 0.000073)
D20160505	-3.350941 ± 0.000347	(1 ± 0.000104)

D20160628	-2.990824 ± 0.000536	(1 ± 0.000179)
D20160922	-4.001674 ± 0.000123	(1 ± 0.000031)
D20161108	-4.451140 ± 0.000078	(1 ± 0.000018)



Summary of RX/WS Ratio (Corrected for OE):


Mean value:	-4.066139
Standard deviation:	0.272486
Standard Error:	0.102747
Relative Standard Error:	0.025269

Power Imbalance

List of Measurements

D20150804	1.003523 ± 0.000051	(1 ± 0.000051)
D20150827	1.003430 ± 0.000040	(1 ± 0.000040)
D20151013	1.001645 ± 0.000043	(1 ± 0.000043)
D20160108	0.990394 ± 0.000045	(1 ± 0.000045)

<i>D20160505</i>	<i>0.999665 ± 0.000048</i>	<i>(1 ± 0.000048)</i>
<i>D20160628</i>	<i>1.002389 ± 0.000046</i>	<i>(1 ± 0.000046)</i>
<i>D20160922</i>	<i>0.996618 ± 0.000047</i>	<i>(1 ± 0.000047)</i>
<i>D20161108</i>	<i>0.998610 ± 0.000049</i>	<i>(1 ± 0.000049)</i>



Summary of Power Imbalance:

<i>Mean value:</i>	<i>0.999601</i>
<i>Standard deviation:</i>	<i>0.004492</i>
<i>Standard Error:</i>	<i>0.001694</i>
<i>Relative Standard Error:</i>	<i>0.001695</i>

Summary

<i>Description</i>	<i>Value</i>	<i>Std Dev</i>	<i>Std Err</i>	<i>Rel Err:</i>
<i>OE (e)</i>	<i>0.9274</i>	<i>0.1095</i>	<i>0.0413</i>	<i>0.0445</i>
<i>TX/WS (a1a2)</i>	<i>-3.3175</i>	<i>0.1979</i>	<i>0.0746</i>	<i>0.0225</i>
<i>RX/WS (b1b2)</i>	<i>-4.0661</i>	<i>0.2725</i>	<i>0.1027</i>	<i>0.0253</i>
<i>W (a5)</i>	<i>1.0000</i>	<i>----</i>	<i>----</i>	<i>0.0209</i>

Published with MATLAB® 8.0