
LHOY-End RxPD and TxPD Calibration Trends

Table of Contents

About	1
WS/TX Ratio when WS is at TX (Inner Beam)	2
WS/TX Ratio when WS is at TX (Outer Beam)	3
WS/TX Ratio when WS is at RX (Inner Beam)	4
WS/TX Ratio when WS is at RX (Outer Beam)	6
RX/TX Ratio (Inner Beam)	7
RX/TX Ratio (Outer Beam)	8
Optical Efficiency of Inner Beam	10
Optical Efficiency of Outer Beam	11
Total Optical Efficiency	12
TX/WS Ratio	14
RX/WS Ratio	15
TX/WS Ratio (Corrected for OE)	16
RX/WS Ratio (Corrected for OE)	18
Power Imbalance	19
Summary	20

About

This document contains the Pcal PD (TxPD and RxPD) Calibration trends. The first six sections contain the six ratios measured at the end-station labeled as m1, m2m6. The section that follows contains the relevant information calculated from these measurements which include Optical Efficiency, Power Imbalance, TX/WS and RX/WS ratio.

Understanding Each Section

Each section contains a list of measurements with Magnitude, Standard Error (Std Err) and Relative Error (Rel Err) for each measurement. The list is followed by two plot figures with Magnitude on the first plot and the the Normalized Magnitude on the second. Each section ends with a summary that contains the weighted mean of all the measurement along with their Standard Deviation (Std Dev), Std Err and Rel Err where each of


$$\begin{aligned} \text{Mean} &= \text{sum}(x(i)*w(i))/\text{sum}(w(i)) \\ \text{Std Dev} &= \text{sqr}t(\text{sum}(w(i)*(x(i)-x_mean)^2)/((n-1)/n*\text{sum}(w(i)))) \\ \text{Std Err} &= \text{Std Dev}/\text{sqr}t(n) \\ \text{Rel Err} &= \text{Std Err}/\text{Mean} \end{aligned}$$

Report created on 24-Jan-2017

WS/TX Ratio when WS is at TX (Inner Beam)

List of Measurements

<i>Date</i>	<i>$m1 \pm SE_{\{m1\}}$</i>	<i>Normalized</i>
<i>D20150811</i>	<i>-0.180846 ± 0.000008</i>	<i>(1 ± 0.000044)</i>
<i>D20150827</i>	<i>-0.180614 ± 0.000007</i>	<i>(1 ± 0.000037)</i>
<i>D20151013</i>	<i>-0.180474 ± 0.000006</i>	<i>(1 ± 0.000034)</i>
<i>D20151222</i>	<i>-0.180696 ± 0.000008</i>	<i>(1 ± 0.000044)</i>
<i>D20160505</i>	<i>-0.178817 ± 0.000008</i>	<i>(1 ± 0.000043)</i>
<i>D20160628</i>	<i>-0.178884 ± 0.000007</i>	<i>(1 ± 0.000041)</i>
<i>D20160927</i>	<i>-0.179158 ± 0.000008</i>	<i>(1 ± 0.000046)</i>
<i>D20161011</i>	<i>-0.179883 ± 0.000010</i>	<i>(1 ± 0.000054)</i>
<i>D20161031</i>	<i>-0.181175 ± 0.000007</i>	<i>(1 ± 0.000040)</i>
<i>D20161109</i>	<i>-0.162027 ± 0.000016</i>	<i>(1 ± 0.000102)</i>
<i>D20170117</i>	<i>-0.161686 ± 0.000024</i>	<i>(1 ± 0.000146)</i>
<i>D20170124</i>	<i>-0.163074 ± 0.000006</i>	<i>(1 ± 0.000035)</i>




Summary of WS/TX when WS is at TX (Inner beam) [m1]:
Mean value: -0.176915
Standard deviation: 0.007040
Standard Error: 0.002121
Relative Standard Error: 0.011987

WS/TX Ratio when WS is at TX (Outer Beam)

List of Measurements

Date	$m2 \pm SE_{\{m2\}}$	Normalized
D20150811	-0.183815 ± 0.000007	(1 ± 0.000039)
D20150827	-0.183788 ± 0.000006	(1 ± 0.000035)
D20151013	-0.184127 ± 0.000007	(1 ± 0.000036)
D20151222	-0.184502 ± 0.000005	(1 ± 0.000027)
D20160505	-0.183087 ± 0.000008	(1 ± 0.000044)
D20160628	-0.183461 ± 0.000008	(1 ± 0.000042)
D20160927	-0.184007 ± 0.000008	(1 ± 0.000044)

<i>D20161011</i>	-0.181887 ± 0.000010	(1 ± 0.000053)
<i>D20161031</i>	-0.183163 ± 0.000008	(1 ± 0.000041)
<i>D20161109</i>	-0.163651 ± 0.000015	(1 ± 0.000094)
<i>D20170117</i>	-0.165437 ± 0.000024	(1 ± 0.000142)
<i>D20170124</i>	-0.165109 ± 0.000006	(1 ± 0.000033)




Summary of WS/TX when WS is at TX (Outer Beam) [m2]:
Mean value: -0.180448
Standard deviation: 0.007498
Standard Error: 0.002258
Relative Standard Error: 0.012516

WS/TX Ratio when WS is at RX (Inner Beam)

List of Measurements

<i>Date</i>	$m3 \pm SE_{\{m3\}}$	<i>Normalized</i>
<i>D20150811</i>	-0.179002 ± 0.000008	(1 ± 0.000043)

D20150827	-0.176861 ± 0.000006	(1 ± 0.000036)
D20151013	-0.178931 ± 0.000006	(1 ± 0.000035)
D20151222	-0.179098 ± 0.000007	(1 ± 0.000037)
D20160505	-0.177402 ± 0.000008	(1 ± 0.000043)
D20160628	-0.177210 ± 0.000008	(1 ± 0.000043)
D20160927	-0.178050 ± 0.000008	(1 ± 0.000046)
D20161011	-0.177663 ± 0.000009	(1 ± 0.000050)
D20161031	-0.179651 ± 0.000008	(1 ± 0.000044)
D20161109	-0.160833 ± 0.000019	(1 ± 0.000120)
D20170117	-0.157927 ± 0.000018	(1 ± 0.000114)
D20170124	-0.161589 ± 0.000006	(1 ± 0.000038)




Summary of WS/TX when WS is at RX (Inner Beam) [m3]:
Mean value: -0.175460

Standard deviation: 0.006641
Standard Error: 0.002001
Relative Standard Error: 0.011401

WS/TX Ratio when WS is at RX (Outer Beam)

List of Measurements

<i>Date</i>	<i>$m4 \pm SE_{\{m4\}}$</i>	<i>Normalized</i>
<i>D20150811</i>	<i>-0.181381 $\pm$ 0.000008</i>	<i>(1 $\pm$ 0.000046)</i>
<i>D20150827</i>	<i>-0.181640 $\pm$ 0.000006</i>	<i>(1 $\pm$ 0.000031)</i>
<i>D20151013</i>	<i>-0.181781 $\pm$ 0.000006</i>	<i>(1 $\pm$ 0.000034)</i>
<i>D20151222</i>	<i>-0.182083 $\pm$ 0.000006</i>	<i>(1 $\pm$ 0.000031)</i>
<i>D20160505</i>	<i>-0.180668 $\pm$ 0.000007</i>	<i>(1 $\pm$ 0.000040)</i>
<i>D20160628</i>	<i>-0.180940 $\pm$ 0.000007</i>	<i>(1 $\pm$ 0.000040)</i>
<i>D20160927</i>	<i>-0.182118 $\pm$ 0.000009</i>	<i>(1 $\pm$ 0.000048)</i>
<i>D20161011</i>	<i>-0.179869 $\pm$ 0.000008</i>	<i>(1 $\pm$ 0.000046)</i>
<i>D20161031</i>	<i>-0.180443 $\pm$ 0.000008</i>	<i>(1 $\pm$ 0.000045)</i>
<i>D20161109</i>	<i>-0.161159 $\pm$ 0.000019</i>	<i>(1 $\pm$ 0.000118)</i>
<i>D20170117</i>	<i>-0.159364 $\pm$ 0.000157</i>	<i>(1 $\pm$ 0.000983)</i>
<i>D20170124</i>	<i>-0.163060 $\pm$ 0.000006</i>	<i>(1 $\pm$ 0.000036)</i>




Summary of WS/TX when WS is at RX (Outer Beam) [m4]:
Mean value: -0.178666
Standard deviation: 0.006827
Standard Error: 0.002056
Relative Standard Error: 0.011510

RX/TX Ratio (Inner Beam)

List of Measurements

Date	$m5 \pm SE_{\{m5\}}$	Normalized
D20150811	0.716408 ± 0.000003	(1 ± 0.000005)
D20150827	0.715858 ± 0.000003	(1 ± 0.000004)
D20151013	0.715399 ± 0.000002	(1 ± 0.000003)
D20151222	0.716771 ± 0.000002	(1 ± 0.000003)
D20160505	0.717343 ± 0.000002	(1 ± 0.000003)
D20160628	0.715799 ± 0.000002	(1 ± 0.000003)
D20160927	0.715156 ± 0.000003	(1 ± 0.000004)

<i>D20161011</i>	<i>0.715926 ± 0.000004</i>	<i>(1 ± 0.000006)</i>
<i>D20161031</i>	<i>0.721892 ± 0.000003</i>	<i>(1 ± 0.000004)</i>
<i>D20161109</i>	<i>0.722016 ± 0.000003</i>	<i>(1 ± 0.000004)</i>
<i>D20170117</i>	<i>0.716968 ± 0.000002</i>	<i>(1 ± 0.000003)</i>
<i>D20170124</i>	<i>0.720691 ± 0.000002</i>	<i>(1 ± 0.000003)</i>



Summary of RX/TX (Inner Beam) [m5]:


<i>Mean value:</i>	<i>0.717490</i>
<i>Standard deviation:</i>	<i>0.002418</i>
<i>Standard Error:</i>	<i>0.000728</i>
<i>Relative Standard Error:</i>	<i>0.001015</i>

RX/TX Ratio (Outer Beam)

List of Measurements

<i>Date</i>	<i>m6 ± SE_{m6}</i>	<i>Normalized</i>
<i>D20150811</i>	<i>0.726657 ± 0.000003</i>	<i>(1 ± 0.000004)</i>

D20150827	0.726468 ± 0.000002	(1 ± 0.000003)
D20151013	0.727582 ± 0.000002	(1 ± 0.000003)
D20151222	0.728972 ± 0.000002	(1 ± 0.000003)
D20160505	0.730960 ± 0.000002	(1 ± 0.000003)
D20160628	0.732089 ± 0.000003	(1 ± 0.000004)
D20160927	0.732393 ± 0.000002	(1 ± 0.000003)
D20161011	0.726292 ± 0.000003	(1 ± 0.000004)
D20161031	0.725420 ± 0.000002	(1 ± 0.000003)
D20161109	0.724600 ± 0.000002	(1 ± 0.000003)
D20170117	0.731319 ± 0.000005	(1 ± 0.000007)
D20170124	0.727754 ± 0.000002	(1 ± 0.000003)




Summary of RX/TX (Outer Beam) [m6]:
Mean value: 0.728340

Standard deviation: 0.002598
Standard Error: 0.000783
Relative Standard Error: 0.001074

Optical Efficiency of Inner Beam

List of Measurements

<i>Date</i>	<i>$e_i \pm SE_{\{e_i\}}$</i>	<i>Normalized</i>
<i>D20150811</i>	<i>0.989808 ± 0.000062</i>	<i>(1 ± 0.000062)</i>
<i>D20150827</i>	<i>0.979224 ± 0.000052</i>	<i>(1 ± 0.000053)</i>
<i>D20151013</i>	<i>0.991448 ± 0.000049</i>	<i>(1 ± 0.000049)</i>
<i>D20151222</i>	<i>0.991154 ± 0.000057</i>	<i>(1 ± 0.000058)</i>
<i>D20160505</i>	<i>0.992086 ± 0.000061</i>	<i>(1 ± 0.000061)</i>
<i>D20160628</i>	<i>0.990643 ± 0.000059</i>	<i>(1 ± 0.000060)</i>
<i>D20160927</i>	<i>0.993815 ± 0.000065</i>	<i>(1 ± 0.000065)</i>
<i>D20161011</i>	<i>0.987658 ± 0.000074</i>	<i>(1 ± 0.000075)</i>
<i>D20161031</i>	<i>0.991589 ± 0.000060</i>	<i>(1 ± 0.000060)</i>
<i>D20161109</i>	<i>0.992628 ± 0.000157</i>	<i>(1 ± 0.000158)</i>
<i>D20170117</i>	<i>0.976747 ± 0.000185</i>	<i>(1 ± 0.000190)</i>
<i>D20170124</i>	<i>0.990892 ± 0.000051</i>	<i>(1 ± 0.000052)</i>



Summary of Optical Efficiency (Inner Beam):


Mean value: 0.989551
Standard deviation: 0.004468
Standard Error: 0.001346
Relative Standard Error: 0.001360

Optical Efficiency of Outer Beam

List of Measurements

Date	$e_o \pm SE_{\{e_o\}}$	Normalized
D20150811	0.986759 ± 0.000061	(1 ± 0.000062)
D20150827	0.988317 ± 0.000047	(1 ± 0.000047)
D20151013	0.987256 ± 0.000049	(1 ± 0.000050)
D20151222	0.986890 ± 0.000041	(1 ± 0.000041)
D20160505	0.986788 ± 0.000060	(1 ± 0.000061)
D20160628	0.986257 ± 0.000058	(1 ± 0.000059)
D20160927	0.989731 ± 0.000065	(1 ± 0.000066)

<i>D20161011</i>	<i>0.988909 ± 0.000070</i>	<i>(1 ± 0.000071)</i>
<i>D20161031</i>	<i>0.985153 ± 0.000061</i>	<i>(1 ± 0.000062)</i>
<i>D20161109</i>	<i>0.984771 ± 0.000150</i>	<i>(1 ± 0.000153)</i>
<i>D20170117</i>	<i>0.963290 ± 0.000993</i>	<i>(1 ± 0.001031)</i>
<i>D20170124</i>	<i>0.987593 ± 0.000049</i>	<i>(1 ± 0.000050)</i>



Summary of Optical Efficiency (Outer Beam):


<i>Mean value:</i>	<i>0.987273</i>
<i>Standard deviation:</i>	<i>0.001262</i>
<i>Standard Error:</i>	<i>0.000380</i>
<i>Relative Standard Error:</i>	<i>0.000385</i>

Total Optical Efficiency

List of Measurements

<i>Date</i>	<i>e ± SE_{e}</i>	<i>Normalized</i>
<i>D20150811</i>	<i>0.988271 ± 0.000043</i>	<i>(1 ± 0.000043)</i>

<i>D20150827</i>	<i>0.983810 ± 0.000034</i>	<i>(1 ± 0.000035)</i>
<i>D20151013</i>	<i>0.989331 ± 0.000034</i>	<i>(1 ± 0.000035)</i>
<i>D20151222</i>	<i>0.989000 ± 0.000035</i>	<i>(1 ± 0.000035)</i>
<i>D20160505</i>	<i>0.989405 ± 0.000042</i>	<i>(1 ± 0.000043)</i>
<i>D20160628</i>	<i>0.988423 ± 0.000041</i>	<i>(1 ± 0.000041)</i>
<i>D20160927</i>	<i>0.991746 ± 0.000046</i>	<i>(1 ± 0.000046)</i>
<i>D20161011</i>	<i>0.988287 ± 0.000050</i>	<i>(1 ± 0.000051)</i>
<i>D20161031</i>	<i>0.988353 ± 0.000042</i>	<i>(1 ± 0.000043)</i>
<i>D20161109</i>	<i>0.988680 ± 0.000108</i>	<i>(1 ± 0.000109)</i>
<i>D20170117</i>	<i>0.969941 ± 0.000492</i>	<i>(1 ± 0.000507)</i>
<i>D20170124</i>	<i>0.989232 ± 0.000035</i>	<i>(1 ± 0.000036)</i>



Summary of Total Optical Efficiency:


Mean value: 0.988408

Standard deviation: 0.002114
Standard Error: 0.000637
Relative Standard Error: 0.000644

TX/WS Ratio

List of Measurements

<i>Date</i>	<i>R_TW ± SE_{R_TW}</i>	<i>Normalized</i>
D20150811	-2.742279 ± 0.000011	(1 ± 0.000004)
D20150827	-2.744228 ± 0.000009	(1 ± 0.000003)
D20151013	-2.742723 ± 0.000009	(1 ± 0.000003)
D20151222	-2.738239 ± 0.000009	(1 ± 0.000003)
D20160505	-2.763161 ± 0.000011	(1 ± 0.000004)
D20160628	-2.759803 ± 0.000011	(1 ± 0.000004)
D20160927	-2.753569 ± 0.000011	(1 ± 0.000004)
D20161011	-2.764191 ± 0.000014	(1 ± 0.000005)
D20161031	-2.744705 ± 0.000010	(1 ± 0.000004)
D20161109	-3.070510 ± 0.000023	(1 ± 0.000007)
D20170117	-3.056948 ± 0.000033	(1 ± 0.000011)
D20170124	-3.047086 ± 0.000008	(1 ± 0.000003)



Summary of TX/WS Ratio:


Mean value: -2.804591
Standard deviation: 0.122691
Standard Error: 0.036956
Relative Standard Error: 0.013177

RX/WS Ratio

List of Measurements

Date	$R_{RW} \pm SE_{\{R_{RW}\}}$	Normalized
D20150811	-4.004239 ± 0.000126	(1 ± 0.000032)
D20150827	-4.023526 ± 0.000096	(1 ± 0.000024)
D20151013	-4.000359 ± 0.000097	(1 ± 0.000024)
D20151222	-4.002816 ± 0.000096	(1 ± 0.000024)
D20160505	-4.044736 ± 0.000119	(1 ± 0.000029)
D20160628	-4.042655 ± 0.000118	(1 ± 0.000029)
D20160927	-4.019073 ± 0.000135	(1 ± 0.000033)

<i>D20161011</i>	<i>-4.033790 ± 0.000139</i>	<i>(1 ± 0.000034)</i>
<i>D20161031</i>	<i>-4.019253 ± 0.000127</i>	<i>(1 ± 0.000032)</i>
<i>D20161109</i>	<i>-4.492703 ± 0.000378</i>	<i>(1 ± 0.000084)</i>
<i>D20170117</i>	<i>-4.564432 ± 0.002271</i>	<i>(1 ± 0.000497)</i>
<i>D20170124</i>	<i>-4.461570 ± 0.000117</i>	<i>(1 ± 0.000026)</i>



Summary of RX/WS Ratio:


<i>Mean value:</i>	<i>-4.065211</i>
<i>Standard deviation:</i>	<i>0.142545</i>
<i>Standard Error:</i>	<i>0.042937</i>
<i>Relative Standard Error:</i>	<i>0.010562</i>

TX/WS Ratio (Corrected for OE)

List of Measurements

<i>Date</i>	<i>R_Twc ± SE_{R_Twc}</i>	<i>Normalized</i>
<i>D20150811</i>	<i>-2.758456 ± 0.000060</i>	<i>(1 ± 0.000022)</i>

D20150827	-2.766624 ± 0.000048	(1 ± 0.000017)
D20151013	-2.757433 ± 0.000048	(1 ± 0.000017)
D20151222	-2.753383 ± 0.000048	(1 ± 0.000018)
D20160505	-2.777877 ± 0.000059	(1 ± 0.000021)
D20160628	-2.775872 ± 0.000058	(1 ± 0.000021)
D20160927	-2.764980 ± 0.000064	(1 ± 0.000023)
D20161011	-2.780474 ± 0.000071	(1 ± 0.000025)
D20161031	-2.760782 ± 0.000059	(1 ± 0.000021)
D20161109	-3.087988 ± 0.000168	(1 ± 0.000054)
D20170117	-3.103593 ± 0.000776	(1 ± 0.000250)
D20170124	-3.063579 ± 0.000054	(1 ± 0.000018)




Summary of TX/WS Ratio (Corrected for OE):
Mean value: -2.799204

Standard deviation: 0.100524
Standard Error: 0.030279
Relative Standard Error: 0.010817

RX/WS Ratio (Corrected for OE)

List of Measurements

<i>Date</i>	<i>R_RWc ± SE_{R_RWc}</i>	<i>Normalized</i>
D20150811	-3.980617 ± 0.000087	(1 ± 0.000022)
D20150827	-3.990691 ± 0.000070	(1 ± 0.000018)
D20151013	-3.978905 ± 0.000069	(1 ± 0.000017)
D20151222	-3.980678 ± 0.000070	(1 ± 0.000018)
D20160505	-4.023196 ± 0.000086	(1 ± 0.000021)
D20160628	-4.019117 ± 0.000084	(1 ± 0.000021)
D20160927	-4.002417 ± 0.000092	(1 ± 0.000023)
D20161011	-4.010027 ± 0.000103	(1 ± 0.000026)
D20161031	-3.995710 ± 0.000086	(1 ± 0.000022)
D20161109	-4.467129 ± 0.000243	(1 ± 0.000054)
D20170117	-4.494785 ± 0.001126	(1 ± 0.000251)
D20170124	-4.437419 ± 0.000080	(1 ± 0.000018)



Summary of RX/WS Ratio (Corrected for OE):


Mean value: -4.046143
 Standard deviation: 0.148577
 Standard Error: 0.044753
 Relative Standard Error: 0.011061

Power Imbalance

List of Measurements

D20150811	0.983938 ± 0.000059	(1 ± 0.000059)
D20150827	0.982844 ± 0.000050	(1 ± 0.000051)
D20151013	0.980135 ± 0.000048	(1 ± 0.000049)
D20151222	0.979459 ± 0.000051	(1 ± 0.000052)
D20160505	0.976585 ± 0.000060	(1 ± 0.000062)
D20160628	0.974999 ± 0.000057	(1 ± 0.000059)
D20160927	0.973728 ± 0.000061	(1 ± 0.000063)
D20161011	0.989087 ± 0.000075	(1 ± 0.000075)

<i>D20161031</i>	<i>0.989175 ± 0.000057</i>	<i>(1 ± 0.000058)</i>
<i>D20161109</i>	<i>0.990030 ± 0.000137</i>	<i>(1 ± 0.000138)</i>
<i>D20170117</i>	<i>0.977093 ± 0.000199</i>	<i>(1 ± 0.000204)</i>
<i>D20170124</i>	<i>0.987783 ± 0.000048</i>	<i>(1 ± 0.000048)</i>



Summary of Power Imbalance:

Mean value:	0.981902
Standard deviation:	0.005408
Standard Error:	0.001629
Relative Standard Error:	0.001659

Summary

Description	Value	Std Dev	Std Err	Rel Err:
OE (e)	0.9884	0.0021	0.0006	0.0006
TX/WS (a1a2)	-2.7992	0.1005	0.0303	0.0108
RX/WS (b1b2)	-4.0461	0.1486	0.0448	0.0111
W (a5)	1.0000	----	----	0.0033

Published with MATLAB® 8.0