
MEAN
Trend from 17-11-13-16-38-02 to 17-11-14-00-38-01
Ch 1: H0:VAC-LY_Y3_PT114B_PRESS_TORR
1.4e-08
1.35e-08
1.3e-08
1.25e-08
1.2e-08
no_conv.
2017-11-13 17:38:25
2017-11-13 19:38:25
2017-11-13 21:38:25
2017-11-13 23:38:25
Removed ISO blank from adapter on 10" valve (valve closed since 1998)
No percieved pressure difference, i.e. seemed already at atmosphere
Pumped down newly installed reducing cross (replaces ISO adapter)
Vented reducing cross
Pumped reducing cross